

SAMPLE

検査成績表

<TEST REPORT>

品名<Model Name> (アナログメグオームハイテスタ<ANALOG MΩ HiTESTER>)
形名<Model Number> (IR4032) *1
製造番号<Serial No.> (No.180523456)
検査年月日<Test Date> (2018-05-09)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (23.0 °C, 60 %rh)

項目 <Item>	レンジ <Range>	計器指示値 <Indicated Value>	許容範囲 <Tolerance>	校正値 <Calibration Value>
確度(許容差) <Accuracy (Limit Deviation Tolerance)>				
125V		0.01 MΩ	0.0090 MΩ ~ 0.0110 MΩ	(0.0100 MΩ)
		0.1 MΩ	0.0950 MΩ ~ 0.1050 MΩ	(0.1000 MΩ)
		2 MΩ	1.90 MΩ ~ 2.10 MΩ	(2.00 MΩ)
		50 MΩ	47.5 MΩ ~ 52.5 MΩ	(50.0 MΩ)
		100 MΩ	90 MΩ ~ 110 MΩ	(100 MΩ)
250V		0.01 MΩ	0.0090 MΩ ~ 0.0110 MΩ	(0.0100 MΩ)
		0.1 MΩ	0.0950 MΩ ~ 0.1050 MΩ	(0.1000 MΩ)
		2 MΩ	1.90 MΩ ~ 2.10 MΩ	(2.00 MΩ)
		50 MΩ	47.5 MΩ ~ 52.5 MΩ	(50.0 MΩ)
		100 MΩ	90 MΩ ~ 110 MΩ	(100 MΩ)
500V		0.01 MΩ	0.0090 MΩ ~ 0.0110 MΩ	(0.0100 MΩ)
		0.1 MΩ	0.0950 MΩ ~ 0.1050 MΩ	(0.1000 MΩ)
		2 MΩ	1.90 MΩ ~ 2.10 MΩ	(2.00 MΩ)
		50 MΩ	47.5 MΩ ~ 52.5 MΩ	(50.0 MΩ)
		100 MΩ	90 MΩ ~ 110 MΩ	(100 MΩ)
ACV (55Hz)		100 V	70 V ~ 130 V	(100 V)
		500 V	470 V ~ 530 V	(500 V)

備考<Note>

・ FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

・ *1 形名欄の記載形名は、製品本体表記に基づいています。

<The model number stated in this document is based on the model number indicated on the instrument.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()