

品名<Model Name> (3CH 電流ユニット<3CH CURRENT UNIT>)
 形名<Model Number> (U8977)
 製造番号<Serial No.> (No.170212345)
 検査年月日<Test Date> (2017-02-22)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (23.0 °C, 50 %rh)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	校正値 <Calibration Value>
1.測定精度<Measurement Accuracy>				
500A定格モード(500A/2V)				
40 A f.s.	-160 mV	-40.120 A ~ -39.880 A	CH1(	-40.000 A)
			CH2(	-40.000 A)
			CH3(	-40.000 A)
	-80 mV	-20.120 A ~ -19.880 A	CH1(	-20.000 A)
			CH2(	-20.000 A)
			CH3(	-20.000 A)
	0 V	-0.120 A ~ 0.120 A	CH1(	0.000 A)
			CH2(	0.000 A)
			CH3(	0.000 A)
	80 mV	19.880 A ~ 20.120 A	CH1(	20.000 A)
			CH2(	20.000 A)
			CH3(	20.000 A)
	160 mV	39.880 A ~ 40.120 A	CH1(	40.000 A)
			CH2(	40.000 A)
			CH3(	40.000 A)
100 A f.s.	400 mV	99.70 A ~ 100.30 A	CH1(	100.00 A)
			CH2(	100.00 A)
			CH3(	100.00 A)
200 A f.s.	800 mV	199.40 A ~ 200.60 A	CH1(	200.00 A)
			CH2(	200.00 A)
			CH3(	200.00 A)
400 A f.s.	1.6 V	398.80 A ~ 401.20 A	CH1(	400.00 A)
			CH2(	400.00 A)
			CH3(	400.00 A)
1000 A f.s.	4 V	997.0 A ~ 1003.0 A	CH1(	1000.0 A)
			CH2(	1000.0 A)
			CH3(	1000.0 A)
2000 V f.s.	8 V	1994.0 A ~ 2006.0 A	CH1(	2000.0 A)
			CH2(	2000.0 A)
			CH3(	2000.0 A)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E.E. CORPORATION

2/2

品名<Model Name> (3CH 電流ユニット<3CH CURRENT UNIT>)
 形名<Model Number> (U8977)
 製造番号<Serial No.> (No.170212345)
 検査年月日<Test Date> (2017-02-22)
 <YYYY-MM-DD>

項目 ＜Item＞	レンジ ＜Range＞	入力 ＜Input＞	許容範囲 ＜Tolerance＞				校正値 ＜Calibration Value＞			
1.測定確度＜Measurement Accuracy＞										
200A定格モード(200A/2V)										
	20 A f.s.	200 mV	19.940	A	～	20.060	A	CH1(	20.000	A)
								CH2(	20.000	A)
								CH3(	20.000	A)
	40 A f.s.	400 mV	39.880	A	～	40.120	A	CH1(	40.000	A)
								CH2(	40.000	A)
								CH3(	40.000	A)
	100 A f.s.	1 V	99.70	A	～	100.30	A	CH1(	100.00	A)
								CH2(	100.00	A)
								CH3(	100.00	A)
	200 A f.s.	2 V	199.40	A	～	200.60	A	CH1(	200.00	A)
								CH2(	200.00	A)
								CH3(	200.00	A)
	400 A f.s.	-4 V	-401.20	A	～	-398.80	A	CH1(	-400.00	A)
								CH2(	-400.00	A)
								CH3(	-400.00	A)
		2 V	-201.20	A	～	-198.80	A	CH1(	-200.00	A)
								CH2(	-200.00	A)
								CH3(	-200.00	A)
		0 V	-1.20	A	～	1.20	A	CH1(	0.00	A)
								CH2(	0.00	A)
								CH3(	0.00	A)
		2 V	198.80	A	～	201.20	A	CH1(	200.00	A)
								CH2(	200.00	A)
								CH3(	200.00	A)
		4 V	398.80	A	～	401.20	A	CH1(	400.00	A)
								CH2(	400.00	A)
								CH3(	400.00	A)
	1000 A f.s.	10 V	997.0	A	～	1003.0	A	CH1(	1000.0	A)
								CH2(	1000.0	A)
								CH3(	1000.0	A)