

SAMPLE 検査成績表
<TEST REPORT>

1/1

品名<Model Name> ('アナログユニット<ANALOG UNIT>)
 形名<Model Number> (8945)
 製造番号<Serial No.> (No. 091234567)
 検査年月日<Test Date> (2011-11-03)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.0 °C, 50 %rh)

項目 <Item>	レンジ <Range>	試験点 <Test Point>	許容範囲 <Tolerance>	測定結果 <Measured Value>
--------------	----------------	---------------------	---------------------	--------------------------

ゼロポジション確度<Zero Position>

10 mV/DIV 0 mV ± 0.10 mV (0.08 mV)

確度<Accuracy>

10 mV/DIV	100 mV	±	0.25 mV	(99.75 mV)
20 mV/DIV	200 mV	±	0.50 mV	(199.50 mV)
50 mV/DIV	500 mV	±	1.25 mV	(498.75 mV)
100 mV/DIV	1 V	±	0.0025 V	(0.9975 V)
200 mV/DIV	2 V	±	0.0050 V	(1.9950 V)
500 mV/DIV	5 V	±	0.0125 V	(4.9875 V)
1 V/DIV	10 V	±	0.025 V	(9.975 V)
2 V/DIV	20 V	±	0.050 V	(19.950 V)
5 V/DIV	50 V	±	0.125 V	(49.875 V)
10 V/DIV	100 V	±	0.25 V	(99.75 V)
20 V/DIV	200 V	±	0.50 V	(199.50 V)
50 V/DIV	350 V	±	1.25 V	(348.75 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No8945-4